


Xylem Analytics Germany GmbH  
Am Achalaich 11  
D-82362 Weilheim  
Germany



## UK - DECLARATION OF CONFORMITY

We declare under our sole responsibility that the product described herein
TitroLine 7500 KF
is in conformity with the requirements of the following documents:
Electromagnetic Compatibility Regulation 2016 BS EN 61326-1:2013
Electrical Equipment (Safety) Regulation 2016 BS EN 61010-1:2010
The Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulation 2012 BS EN IEC 63000: 2018
Radio Equipment Regulation 2017

Mainz, 22.08.2022  
Conf. No.: UK-Titrat015

  
Robert Reining  
Managing Director



SI Analytics®

-ebro-®



are brands of Xylem Inc. or one of its subsidiaries.