

Xylem Analytics Germany GmbH
Am Achalaich 11
D-82362 Weilheim
Germany



UK - DECLARATION OF CONFORMITY

We declare under our sole responsibility that the products described herein
pH3210, pH3310, pH/ION3310, pH3310IDS Oxi3205, Oxi3210, Oxi3310, Oxi3315, Oxi3310IDS Cond3210, Cond3310, Cond3310IDS Multi3310IDS, Multi3510IDS
are in conformity with the requirements of the following documents:
Electromagnetic Compatibility Regulations 2016 BS EN 61326-1:2013
Electrical Equipment (Safety) Regulations 2016 BS EN 61010-1:2010
The Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012 BS EN IEC 63000:2018

Weilheim, July 1, 2022

Dr. Robert Reining
Managing Director

KONF. No.: UK100a



SI Analytics®



are brands of Xylem Inc. or one of its subsidiaries.