Xylem Analytics Germany GmbH Am Achalaich 11 D-82362 Weilheim Germany



UK - DECLARATION OF CONFORMITY

We declare under our sole responsibility that the products described herein

> pH7310, pH7310P pH/ION7320, pH/ION7320P Oxi7310, Oxi7310P Cond7310, Cond7310P Multi9310, Multi9310P

are in conformity with the requirements of the following documents:

> **Electromagnetic Compatibility** Regulations 2016 BS EN 61326-1:2013

Electrical Equipment (Safety) Regulations 2016 BS EN 61010-1:2010

The Restriction of the Use of Certain Hazardous Substances in Electrical and Electronic Equipment Regulations 2012 BS EN IEC 63000:2018

Weilheim, July 1, 2022

Dr. Robert Reining Managing Director

KONF. No.: UK106a







