

Hersteller-Zertifikat Manufacturer's Certificate

Modell / Model : **PL 7**
Bestell-Nr./ Order No. : **109 120**
Chargen-Nr./ Lot-No. : **04.21-1**

Die oben genannte Pufferlösung entspricht in Bezug auf die wirksamen Puffersubstanzen der DIN Norm 19266.

Die Prüfung der Lösung erfolgt anhand einer Vergleichsmessung mit einer Glas-Einstabmesskette gegen sekundäres Referenzmaterial (Merck Charge HC60268302, rückführbar auf SRM 186-I-g und 186-II-g von NIST) gleichen Nominalwerts.

Zertifizierter Wert (bei $25,0^{\circ}\text{C} \pm 0,1^{\circ}\text{C}$):

$6,865 \pm 0,020$ pH-Einheiten

Die angegebene Unsicherheit entspricht der erweiterten Messunsicherheit U. Diese ergibt sich aus der Standardmessunsicherheit u (Typ B) multipliziert mit dem Erweiterungsfaktor ($k = 2$). Die Standardmessunsicherheit wurde gemäß dem „Guide to the Expression of Uncertainty in Measurement“ (ISO, 1995) ermittelt.

Die Lösung enthält zur Vorbeugung gegen Pilz- oder Bakterienbefall ein Desinfektionsmittel im Volumenanteil von 0,02% (Wirkstoff: Benzalkoniumchlorid).

Die Mindesthaltbarkeit im ungeöffneten Zustand ist auf der Rückseite der Flasche oder seitlich etikettiert und gibt gleichzeitig die vollständige Chargen-Nummer an.

Beispiel :

04.19-3 $\Leftrightarrow$ Haltbar bis April 2019 - Charge 3
im Herstellungsmonat

The above listed buffer solution is complying to DIN standard 19266 in terms of the effective buffer ingredients.

The solution is tested by comparative measurement against a secondary reference buffer solution of the same nominal value (Merck batch HC60268302, traceable to SRM 186-I-g and 186-II-g of NIST) using a glass combined electrode.

Certified Value (at $25,0^{\circ}\text{C} \pm 0,1^{\circ}\text{C}$):

6.865 ± 0.020 pH units

The uncertainty mentioned above represents the expanded uncertainty U, obtained from the standard uncertainty u (type B) by a coverage factor ($k = 2$). The standard uncertainty is calculated according to the „Guide to the Expression of Uncertainty in Measurement“ (ISO, 1995).

The solution contains 0.02% in volume of a disinfectant to prevent befall by fungus or by bacteria (effective constituent: Benzalkonium Chloride).

The recommended time of best use valid for the unopened state is labeled on the backside or on the side of the bottle and represents simultaneously the complete lot number.

Example :

04.19-3 $\Leftrightarrow$ Best use before April 2019 - Lot 3
in month of production

Weilheim, 2018-04-11

Alexander Strauß
Qualitätssicherung / Quality Management