

IQ SENSOR NET Provides Reliable, Accurate Continuous Measurement for Troubleshooting Nitrification and Denitrification

The proper tools are essential to fixing any problem. Over the years, Ohio EPA's Compliance Assistance Unit (CAU) has helped dozens of water resource recovery facilities (WRRF) get back into and maintain compliance with NPDES (National Pollutant Discharge Elimination System) discharge permits. Their field toolkit has included an assortment of batch sampling kits and handheld instruments. Each tool had its place but each also had limitations and, as a result, the compliance puzzle often was missing important pieces.

Increasingly, as nutrient limits were incorporated into discharge permits, the missing piece was characterizing the dynamics of nitrification and denitrification.

A new tool was needed with the following four features:

Continuous

Continuous measurement with datalogging is essential to diagnosing problems. Batch tests on manual samples or laboratory measurements from grab samples provide a snapshot but do not provide insight into treatment dynamics.

Furthermore, many problems occur overnight or on weekends when operators are not present to take measurements.

Simple

The tool needs to be simple and easy to understand.

Many operators are part-time and most have additional responsibilities besides wastewater treatment.

Tools that are overly complicated or require a lot of maintenance take time away from the operator to monitor treatment now and will eventually be neglected eliminating the possibility for greater understanding of the process for the future.

Inexpensive

The tool needs to be inexpensive. It is important that the acquisition cost be comparable with other tools to fit into the budget.

It is also important that ownership cost is low to not significantly increase the annual operating expenses. Furthermore, many wastewater facilities are small facilities serving institutions that also have a limited budget. Even better is a tool that pays for itself by enabling reduced O&M costs for energy or chemical inputs.



Figure 1: IQ SENSOR NET VARIION® sensor in Anoxic Tank



Figure 2: IQ SENSOR NET VARIION® sensor in Final Effluent

Versatile

Wastewater treatment requires balancing a number of concurrent reactions. Tools that enable simultaneous evaluation of multiple parameters reduce the time required for troubleshooting and improve the result are highly desirable. The IQ SENSOR NET of the WTW brand has all the desired features and offered a reliable option for monitoring ammonium and nitrate with one instrument. The CAU equipment budget was limited so they needed proof of performance before placing an order.



Xylem Analytics provided a trial system for them to evaluate and helped set up the system at the Millersburg, Ohio wastewater treatment plant. Two VARiON® sensors were deployed, one in the anoxic basin (Figure 1) and the other in the final effluent channel (Figure 2). Each digital sensor was outfitted with ion selective electrodes (ISE) for ammonium, nitrate, and potassium in addition to a reference electrode and a temperature probe. An MIQ/Blue PS bluetooth module was utilized to connect the two measurement locations on the same network without the need for additional cabling. Data retrieved from the Millersburg trial demonstrates the accuracy of the IQ SENSOR NET measuring system. A comparison of ammonium measurements recorded by the VARiON® sensor and daily reference measurements made with a portable colorimeter from grab samples taken from the anoxic tank is shown in Figure 3. There is close agreement between the online measurements and the grab measurements in most cases. There was no attempt to determine which value was more correct during times that the values were relatively far apart.




Figure 3: Ammonia Nitrogen Concentration in the Anoxic Basin with IQ SENSOR NET and Lab Colorimeter Samples

VARiON® measurements from the effluent location are compared with measurements from the existing cabinet analyzer and grab sample measurements in Figure 4.




Figure 4: Ammonia Nitrogen Concentration in the Final Effluent with IQ SENSOR NET, Online Cabinet Analyzer, and Lab Colorimeter Samples

The VARiON® and cabinet analyzer values trended closely although the VARiON® reported values were generally higher by 0.5 to 2.0 mg/L. Laboratory analysis of grab measurements did not provide conclusive evidence that one method or the other is more accurate. Colorimeter measurements sometimes were in closer agreement with the VARiON® and other times in closer agreement with the cabinet analyzer. The IQ SENSOR NET proved its value to the OEPA CAU for monitoring nitrification and denitrification.

Furthermore, maintenance also proved to be simple. A matrix adjustment was required for the ammonium, nitrate, and potassium electrodes at commissioning but the measuring accuracy remained stable thereafter without further adjustment. Fouling of the sensors will occur in wastewater environments. Some accumulation was evident on the sensors used in this project (Figure 5) but the important observation was that the reliability of the measurement was not affected, even after six weeks continuous operation.




Figure 5: VARiON® sensor After 6 Weeks of Continuous Operation in an Anoxic Bioreactor

The results of the trial led the CAU to order two VARiON® sensors, a 2020XT controller, and a Bluetooth set for simple set-up and communication at multiple measuring locations. The CAU has since deployed their system at multiple locations for troubleshooting nitrification/denitrification. The online measurements quickly identify and confirm problems saving time and increasing productivity of the CAU.

Do you have further questions?
Please contact our
Customer Care Center:

Xylem Analytics Germany Sales
GmbH & Co. KG, WTW
Am Achalaich 11
82362 Weilheim, Germany
Phone +49 881 1830
Fax +49 881 183-420
Info.WTW@xylem.com