

Hersteller-Zertifikat Manufacturer's Certificate

Modell / Model : **PL 12 / PLE 12**
Bestell-Nr./ Order No. : **109 400 / 109 401**
Chargen-Nr./ Lot-No. : **07.23-1**

Die oben genannte Pufferlösung entspricht in Bezug auf die wirksamen Puffersubstanzen der DIN Norm 19266.

Die Prüfung der Lösung erfolgt anhand einer Vergleichsmessung mit einer Glas-Einstabmesskette gegen DAkkS-zertifiziertes rückführbares Referenzmaterial gleichen Nominalwerts (Merck Charge 172111C).

Zertifizierter Wert (bei $25,0^{\circ}\text{C} \pm 0,1^{\circ}\text{C}$):

$12,47 \pm 0,05$ pH-Einheiten

Die angegebene Unsicherheit entspricht der erweiterten Messunsicherheit U. Diese ergibt sich aus der Standardmessunsicherheit u (Typ B) multipliziert mit dem Erweiterungsfaktor ($k = 2$). Die Standardmessunsicherheit wurde gemäß dem „Guide to the Expression of Uncertainty in Measurement“ (ISO, 1995) ermittelt.

Die Lösung enthält zur Vorbeugung gegen Pilz- oder Bakterienbefall ein Desinfektionsmittel im Volumenanteil von 0,02% (Wirkstoff: Benzalkoniumchlorid).

Die Mindesthaltbarkeit im ungeöffneten Zustand ist auf der Rückseite der Flasche oder seitlich etikettiert und gibt gleichzeitig die vollständige Chargen-Nummer an.

Beispiel :

04.24-3 $\Leftrightarrow$ Haltbar bis April 2024 - Charge 3
im Herstellungsmonat

The above listed buffer solution is complying to DIN standard 19266 in terms of the effective buffer ingredients.

The solution is tested by comparative measurement against a DAkkS certified traceable reference buffer solution of the same nominal value (Merck batch 172111C) using a glass combined electrode.

Certified Value (at $25,0^{\circ}\text{C} \pm 0,1^{\circ}\text{C}$):

12.47 ± 0.05 pH units

The uncertainty mentioned above represents the expanded uncertainty U, obtained from the standard uncertainty u (type B) by a coverage factor ($k = 2$). The standard uncertainty is calculated according to the „Guide to the Expression of Uncertainty in Measurement“ (ISO, 1995).

The solution contains 0.02% in volume of a disinfectant to prevent befall by fungus or by bacteria (effective constituent: Benzalkonium Chloride).

The recommended time of best use valid for the unopened state is labeled on the backside or on the side of the bottle and represents simultaneously the complete lot number.

Example :

04.24-3 $\Leftrightarrow$ Best use before April 2024 - Lot 3
in month of production

Weilheim, 2021-07-09

S. Dreßler
Qualitätssicherung / Quality Management